

THE FOLLOWING PAGE CONTAINS THE CONNECTION DETAILS

It is your responsibility to ensure that the transducer identification matches the connection details, including any RDP modification (T.M) references.

Please ensure that the transducer has not been modified or fitted with a cable or extension.


Particularly on DC-DC powered transducers, incorrect connection may cause **IRREPERABLE DAMAGE** to the transducer which is **NOT COVERED BY WARRANTY**

IF YOU ARE IN ANY DOUBT PLEASE CONTACT US



DCW SERIES

TM321A & B STD AND OPT 7 CABLES



	-FS	0	+FS
TM321A	4mA	12mA	20mA
TM321B	20mA	12mA	4mA

See calibration certificate for exact figures.

BANDWIDTH 250Hz (flat)
NOISE 50uA peak to peak max
ZERO TEMPCO 0.01% F.S./°C (Typical)
SPAN TEMPCO 0.03% F.S./°C (Typical)
OPERATING TEMP -10°C to +70°C
CALIBRATION 24V dc (see calibration certificate)
SUPPLY Smooth d.c. voltage 12 to 36V

**NOTE : INCORRECT CONNECTION MAY CAUSE IRREPARABLE DAMAGE
 CONTACT OUR SALES DEPARTMENT IF YOU REQUIRE ASSISTANCE**

RDP U.K.
Tel 01902 457512
Fax 01902 452000



Certificate No. FM 13141
 BS EN ISO 9001:2000

**MADE IN THE
 UNITED KINGDOM**

Doc. Ref. WDUK1454
 ISSUE 3 - 12 AUG 2004

RDP U.S.A.
Tel 610 469 0850
Fax 610 469 0852